## Corning Incorporated Life Sciences

Registered ISO 9001:2008

# **Product Description**

Catalog Number: 3357

Product Description: Costar ® 96-well polypropylene, v-bottom plate, without lid, sterile

**Component Materials:** 

Plate - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and

closures.

**Product Dimensions:** 

Length of Plate - 5.037 in. Diameter of Well @ top - .270 in. Width of Plate - 3.365 in. Height - .560 in. Depth to the bottom of - .438 in. Working Volume per well - 200 µL

well

Tolerances of - +/- .010 in. with storage

Dimensions mat sealed

#### Sterilization:

The product has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*Sterility Assurance Level: SAL 10<sup>-3</sup>

### RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

#### Surface Characterization:

Surface is characterized to be hydrophobic and non-charged for solution based assay.

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

# Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling and correct

product configuration.

## Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 11