

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

# Product Description

**Catalog Number:** 3494

**Product Description:** Costar® Transwell-COL®, 12mm, 3.0 µm pore size

**Component Materials:**

- |                |   |  |
|----------------|---|--|
| Plate/Lid      | - | Virgin Polystyrene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |
| Transwell body | - | Virgin Polystyrene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |
| Filter         | - | Transparent collagen-treated PTFE membrane.  |
| Collagen       | - | Types I and III bovine placenta  |

**Product Dimensions:**

Length of Plate	-	5.035 in.	Bottom of Transwell to plate	-	.040 in.
Width of Plate	-	3.370 in.	Volume added / plate	-	1.5 mL
Height with Lid	-	.890 in.	Volume added / Transwell	-	0.5 mL
Tolerances of Dimensions	-	+/- .050 in.			

**Sterilization:**

The Transwells are sterilized according to ANSI/AAMI/ISO 11135 *Medical Devices - Validation and routine control of ethylene oxide sterilization* and released after successful sterility testing of Biological Indicators. Sterility Assurance Level: SAL 10<sup>-6</sup> The plates are irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*. Sterility Assurance Level: SAL 10<sup>-3</sup>

**Cell Attachment and Growth Characteristics:**

The lot has been tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line in a serum supplemented media.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:	Visual and microscopic examination of the product.
Packaging:	Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.
Collagen Assay:	Colorimetric determination of collagen attachment to the membrane.
Cell Attachment/Growth:	Four to five day cell attachment and growth to $\geq 95\%$ confluency.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 5