

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Product Description

Catalog Number: 3622

Product Description: Costar® 1.7 mL Microcentrifuge Tube, rainbow pack

Component Materials:

Tube - Polypropylene with yellow, blue, green, pink and purple colorant, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

| | | | | | |
|------------|---|-----------|--------------------|---|-------|
| Length | - | 1.535" | I.D. @ top of tube | - | .362" |
| Tolerances | - | +/- .050" | O.D. @ cap | - | .500" |

Special Feature:

The tube is autoclavable.

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Pyrogens:

The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987)*. Results: less than 0.1 EU/mL

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 4