Corning Incor Life Scien Registere ISO 9001:20	Product Description
Catalog Number:	431206
Product Description:	Corning ${\rm I}$ 1 Liter filter system, cellulose acetate membrane, 0.45 $\mu m$ pore size, 500 mL funnel
Component Materials: Funnel, Funnel Lid & - Receiver/Storage Bottle	Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and closures.
Funnel Adapter, Hose - Connector & Plug Seal Cap	Virgin High Density Polyethylene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. Heavy metal free (meets <i>CONEG</i> req.) color concentrate.
Membrane -	Cellulose Acetate, 0.45 μm pore size Polyethylene Foam liner Cellulose Acetate
<b>-</b>	

## Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.* Sterility Assurance Level: SAL 10<sup>-5</sup>

## Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72: *Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing.* Results: less than 0.1EU/mL.

## Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

## **Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures. Membrane Integrity Test: Visual Attributes: Packaging: Forward pressurization of the membrane seal to the funnel. Visual examination of the product. Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Revision Date: 03-01-14

Rev No: 4