

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

# Product Description

**Catalog Number:** 4413

**Product Description:** Costar ® 1.2 mL Cluster Tube Strip, in rack, sterile

**Component Materials:**

Tube Strip - Polypropylene  
Rack - Polypropylene

**Product Dimensions:**

Overall strip length	-	2.820 in.	I.D. @ top	-	0.275 in.
Overall tube length	-	1.830 in.	O.D. @ top	-	0.340 in.
Tolerances	-	+/- .050 in.			

**Sterilization:**

The lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization*.  
Sterility Assurance Level: SAL 10<sup>-3</sup>

**RNase/DNase Testing:**

This lot has been tested and is free of any detectable RNase/DNase contamination.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 5