Corning Incorporated Life Sciences <u>Registered</u> ISO 9001:2008		Product Description			
Catalog Number:	4803				
Product Description:	Corning ® l	Iniversal Fit Pipette Tip, 1	- 200 μL, 480 tip rack		
<b>Component Materials:</b> Tip Rack	closures.	Virgin Polypropylene, meets USP, Class VI requirements for plastic containers and closures. Polypropylene with orange colorant, meets USP, Class VI requirements for plastic containers and closures. Heavy metal free (meets CONEG req.) color concentrate. Polypropylene, meets USP, Class VI requirements for plastic containers and closures.			
Base	containers ar				
<b>Product Dimensions:</b> Overall tip length Tolerences	- 1.995" - +/050"	I.D. smal I.D. large		.028" .215"	

# **Special Feature:**

The tips are autoclavable.

### **RNase/DNase Testing:**

This product has been tested and is free of any detectable RNase/DNase contamination.

### Pyrogens:

The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing. Results: less than 0.1 EU/mL

# Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

## Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures. Visual Attributes: Visual examination of the product. Packaging: Accurate labeling, and correct product configuration.

## Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 5