## Corning Incorporated Life Sciences

Registered ISO 9001:2008

## **Product Description**

Catalog Number: 4834

Costar ® Micro Volume Tip-E, .5 - 10 µL 96 tip rack **Product Description:** 

**Component Materials:** 

Tip Polypropylene, meets USP, Class VI requirements for plastic containers and closures. Rack

Polypropylene with white colorant, meets USP, Class VI requirements for plastic

containers and closures.

Lid Polypropylene, meets USP, Class VI requirements for plastic containers and closures.

**Product Dimensions:** 

Overall tip length I.D. small orifice 1.790" .013" Tolerences +/-.050" I.D. large orifice @ angle -.132"

**Special Feature:** 

The entire unit is autoclavable.

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Pyrogens:

The product has been tested and has met the criteria established in the Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Animal Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987).

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Accurate labeling, and correct product configuration. Packaging:

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 4