Corning Incorporated Life Sciences

Registered ISO 9001:2008

Product Description

Catalog Number: 4876

Product Description: Costar ® Transtar disposable reservoir liner, open

Component Materials:

Reservoir - Virgin white polystyrene

Product Dimensions:

Dimensions

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (TIR 33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*Sterility Assurance Level: SAL 10⁻³

Pyrogens:

The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing. Results: less than 0.1 EU/mL

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct

product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification. The entire lot is sterilized as one unit.

Rev No: 4