

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 3284

Product Description: Corning® CellSTACK® - Filling Accessory - 33mm cap with 1/4" ID tubing and 37mm filter with 1.0µm pore Hydrophobic Glass Lamininate

Component Materials:

- Cap - Virgin High Density Polyethylene, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets CONEG requirements) color concentrate.
- Tubing - C-Flex thermoplastic elastomer, meets *USP, Class VI* requirements for Biological Test for Plastics
- Filter - 37mm filter with 1.0µm pore Hydrophobic Glass Lamininate, meets *USP, Class VI* requirements for Biological Test for Plastics

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR 27) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.
Sterility Assurance Level: SAL 10⁻⁶

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 3