## Corning Incorporated Life Sciences

Registered ISO 9001:2008

## **Product Description**

Catalog Number:	3281			
Product Description:	Corning® 33mm Vented Filling Cap with Tubing and 50mm Filter			
Component Materials:				
Сар	Virgin High Density Polyethylene, meets USP, Class VI requirements for plastic containers and closures. Heavy metal free (meets CONEG requirements) color concentrate.			
Liner	50mm Filter - 0.2µm pore nonwettable (hydrophobic) polyvinylidene difluoride (PVDF) membrane.			
Tubing	Polyethylene			
Filter	Emflon II membrane (hydrophobic PVDF) & Housing: Polypropylene			
Product Dimensions: Length -	.375 in.(3/8")	Inner Dimension -	.95mm	

**BSE/TSE** - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

**Sterilization -** Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10<sup>-6</sup>

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below: Visual Inspection - Pass

visual inspection	-	F a 5 5
Packaging Inspection	-	Pass

## Lot Number Designation:

8 Digit Lot Number: First 3 digits – Julian date, start of manufacturing; Next 2 digits – Year of manufacture; Last 3 digits – Batch identification.

Rev No: 1